



ASTECCO., LTD.
Fukuoka, Japan, 811-2207
<http://www.astec-bio.com/global/>
Email: astec-japan@astec-bio.com
Tel: +81929355666

Certificate of Analysis

Product ID : Pasteur Pipette 1mL
Product Code : ATP-01
Lot # : 202407
Expiry date : 2027-7
Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours	MEA.023.417.2024	Pass	09-23-2024
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.023.426.2024	Pass	09-20-2024
Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation	SMA.023.435.2024	Pass	10-04-2024

Tests conducted by Embryotools

Quality Control Department
ASTECCO., LTD.

