



ASTECCO., LTD.  
Fukuoka, Japan, 811-2207  
<http://www.astec-bio.com/global/>  
Email: [astec-japan@astec-bio.com](mailto:astec-japan@astec-bio.com)  
Tel: +81929355666

## Certificate of Analysis

Product ID : OPU Tube  
Product Code : AIT-14  
Lot # : 202411  
Expiry date : 2027-11  
Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours	MEA.024.142.2025	Pass	01-27-2025
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.024.143.2025	Pass	01-24-2025
Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation	SMA.024.144.2025	Pass	01-24-2025

Tests conducted by Embryotools

Quality Control Department  
ASTECCO., LTD.

