

Certificate of Analysis

Product ID : Sperm Collection Cup
Product Code : AIT-SC
Lot # : 202407
Expiry date : 2027-7
Storage Conditions : 3-30°C, Dry, No direct sunlight

Analysis	Ref#	Results	Date of Completion
Mouse Embryo Assay 80% of one cell stage control embryos to develop to expanded blastocyst within 96 hours	MEA.023.416.2024	Pass	09-23-2024
Proven non-toxic by LAL test. Pass criteria <0.03 EU/device	LAL.023.425.2024	Pass	09-20-2024
Sperm Motility Assay: >0.75 sperm motility by 8 hours post incubation and >0.8 sperm survival by 24 hours post incubation	SMA.023.434.2024	Pass	09-26-2024

Tests conducted by Embryotools

Quality Control Department
ASTEC CO., LTD.

